



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya
Email: info@kenas.go.ke Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION


No	Calibration Field	Calibration Object	Measurand	Laboratory Calibration Method	Standards	Range	Calibration and Measurement Capability (CMC) (\pm expanded uncertainty)	Calibration Site
1	Mass	Non automatic weighing instruments	mass	In-house: ASL/QP/018/MASS Reference Methods: a) EA-4/02, b), EURAMET cg 18:2015	Standard masses Class F1 Traceability BS/MET/2/3/88/613 & 606 Class M2 Traceability BS/MET/2/3/92/732	1mg – 5 kg 5kg - 80000 kg	\pm 0.0008 g to 0.095 g \pm 0.008 kg to 2 kg	Various
2	Mass	Mass standards of class M2 and M3	mass	In-house: ASL/QP/020/MASS Reference Methods: a) EA-4/02, b) KS OIML R 111:2004	Standard masses Class F1 Traceability BS/MET/2/3/92/732 Class M1 Traceability BS/MET/2/3/92/730	1g – 200 kg	\pm 0.10 mg – 1000 mg	ASL laboratory

*The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%.

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Accreditation	01-Oct-2020	01-Oct-2020	30-Sep-2024

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past

Approved by:  Date: 15-April-2021

~~KENAS CEO/Authorized Representative~~

